

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111323\
 Data File : VW032933.D
 Acq On : 13 Nov 2023 15:51
 Operator : SY/MD
 Sample : 05326-02DL 20X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E45K3DL

Quant Time: Nov 14 00:19:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 13 23:16:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	402271	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	402293	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	194254	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	153092	44.997	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.000%
7) Chloroethane-d5	1.532	69	129137	45.811	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.620%
11) 1,1-Dichloroethene-d2	2.059	65	67330	45.421	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.840%
21) 2-Butanone-d5	3.783	46	133842	91.775	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.780%
24) Chloroform-d	4.249	84	263073	44.758	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.520%
26) 1,2-Dichloroethane-d4	4.940	65	166778	46.633	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.260%
32) Benzene-d6	4.960	84	514750	46.914	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.820%
36) 1,2-Dichloropropane-d6	5.989	67	153274	48.059	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.120%
41) Toluene-d8	7.246	98	461307	45.602	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.200%
43) trans-1,3-Dichloroprop...	7.551	79	75164	44.394	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.780%
47) 2-Hexanone-d5	8.024	63	111472	95.405	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.410%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	206132	44.121	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.240%
66) 1,2-Dichlorobenzene-d4	11.561	152	189124	50.105	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.220%
Target Compounds						
					Qvalue	
34) Trichloroethene	5.841	95	4151	1.398	ug/L	86
46) Tetrachloroethene	7.905	164	177315	69.414	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111323\
Data File : VV032933.D
Acq On : 13 Nov 2023 15:51
Operator : SY/MD
Sample : 05326-02DL 20X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E45K3DL

Quant Time: Nov 14 00:19:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 13 23:16:30 2023
Response via : Initial Calibration

