

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111323\
 Data File : VV032928.D
 Acq On : 13 Nov 2023 13:32
 Operator : SY/MD
 Sample : VV1113MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK218

Quant Time: Nov 14 00:18:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 13 23:16:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	397455	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	402348	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	197222	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	127161	37.828	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.660%
7) Chloroethane-d5	1.532	69	105066	37.723	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.440%
11) 1,1-Dichloroethene-d2	2.059	65	55004	37.555	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.120%
21) 2-Butanone-d5	3.786	46	98316	68.232	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	68.230%
24) Chloroform-d	4.249	84	209004	35.990	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	71.980%
26) 1,2-Dichloroethane-d4	4.944	65	132021	37.362	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	74.720%
32) Benzene-d6	4.960	84	415649	37.877	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.760%
36) 1,2-Dichloropropane-d6	5.989	67	123338	38.667	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	77.340%
41) Toluene-d8	7.243	98	369498	36.522	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	73.040%#
43) trans-1,3-Dichloroprop...	7.554	79	61136	36.103	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.200%
47) 2-Hexanone-d5	8.024	63	80828	69.169	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	69.170%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	162339	34.743	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	69.480%
66) 1,2-Dichlorobenzene-d4	11.561	152	151176	39.449	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.900%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111323\
Data File : VV032928.D
Acq On : 13 Nov 2023 13:32
Operator : SY/MD
Sample : VV1113MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK218

Quant Time: Nov 14 00:18:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 13 23:16:30 2023
Response via : Initial Calibration

