

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111323\
 Data File : VW032942.D
 Acq On : 13 Nov 2023 19:20
 Operator : SY/MD
 Sample : VIBLK245
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK245

Quant Time: Nov 14 00:21:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 13 23:16:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	396007	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	400387	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	189784	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	152402	45.503	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.000%
7) Chloroethane-d5	1.532	69	130141	46.897	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.800%
11) 1,1-Dichloroethene-d2	2.060	65	69172	47.402	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.800%
21) 2-Butanone-d5	3.783	46	113651	79.163	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.160%
24) Chloroform-d	4.249	84	263482	45.536	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.080%
26) 1,2-Dichloroethane-d4	4.941	65	165593	47.034	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.060%
32) Benzene-d6	4.960	84	519437	47.567	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.140%
36) 1,2-Dichloropropane-d6	5.989	67	152426	48.021	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.040%
41) Toluene-d8	7.243	98	461879	45.876	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.760%
43) trans-1,3-Dichloroprop...	7.551	79	74361	44.128	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.260%
47) 2-Hexanone-d5	8.024	63	96693	83.150	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.150%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	208419	44.823	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.640%
66) 1,2-Dichlorobenzene-d4	11.561	152	189838	51.479	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111323\
Data File : VV032942.D
Acq On : 13 Nov 2023 19:20
Operator : SY/MD
Sample : VIBLK245
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK245

Quant Time: Nov 14 00:21:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 13 23:16:30 2023
Response via : Initial Calibration

