

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111419\
 Data File : VV013627.D
 Acq On : 14 Nov 2019 18:03
 Operator : SY/MD
 Sample : K5855-11
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAQF9

Quant Time: Nov 15 04:02:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 02:40:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	340736	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	350658	5.00	ug/L	-0.02
60) 1,4-Dichlorobenzene-d4	11.29	152	156959	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	113873	5.20	ug/L	-0.03
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.00%
7) Chloroethane-d5	1.58	69	95867	5.49	ug/L	-0.03
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.80%
11) 1,1-Dichloroethene-d2	2.13	63	145837	3.89	ug/L	-0.03
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.80%
20) 2-Butanone-d5	3.95	46	266630	68.03	ug/L	-0.07
Spiked Amount	50.000	Range	40 - 130	Recovery	=	136.06%#
24) Chloroform-d	4.40	84	232730	6.20	ug/L	-0.02
Spiked Amount	5.000	Range	70 - 125	Recovery	=	124.00%
26) 1,2-Dichloroethane-d4	5.08	65	115269	6.29	ug/L	-0.01
Spiked Amount	5.000	Range	70 - 130	Recovery	=	125.80%
32) Benzene-d6	5.09	84	437794	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.60%
36) 1,2-Dichloropropane-d6	6.11	67	140876	4.86	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.20%
41) Toluene-d8	7.36	98	366821	4.28	ug/L	-0.04
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.60%
43) trans-1,3-Dichloropropene-	7.66	79	49184	4.56	ug/L	-0.04
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.20%
46) 2-Hexanone-d5	8.13	63	202278	47.02	ug/L	-0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.04%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	99771	4.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	154750	5.58	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.60%

Target Compounds

					Ovalue
13) Acetone	2.21	43	18358	7.244	ug/L 68
42) Toluene	7.43	91	38743	0.348	ug/L 98
53) m,p-xylene	9.18	106	4598	0.104	ug/L 95
54) o-xylene	9.58	106	5368	0.129	ug/L 84
63) 1,4-Dichlorobenzene	11.31	146	20544	0.396	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111419\
Data File : VV013627.D
Acq On : 14 Nov 2019 18:03
Operator : SY/MD
Sample : K5855-11
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAQF9

Quant Time: Nov 15 04:02:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 02:40:46 2019
Response via : Initial Calibration

