

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111422\
 Data File : VW028929.D
 Acq On : 14 Nov 2022 11:30
 Operator : SY/MD
 Sample : N5590-04DL 4X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBMX5DL

Quant Time: Nov 15 00:03:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 15 00:01:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	265189	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	241459	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	113587	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	62874	3.889	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	77.800%	
7) Chloroethane-d5	1.568	69	63358	4.458	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	89.200%	
11) 1,1-Dichloroethene-d2	2.108	63	94446	3.206	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	64.200%	
20) 2-Butanone-d5	3.883	46	151491	55.977	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	111.960%	
24) Chloroform-d	4.343	84	158628	4.779	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.600%	
26) 1,2-Dichloroethane-d4	5.027	65	71338	4.777	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.600%	
32) Benzene-d6	5.043	84	331975	4.606	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.200%	
36) 1,2-Dichloropropane-d6	6.066	67	94515	4.696	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	94.000%	
41) Toluene-d8	7.310	98	297524	4.586	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.800%	
43) trans-1,3-Dichloroprop...	7.622	79	30471	4.642	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	92.800%	
46) 2-Hexanone-d5	8.085	63	134119	47.870	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.740%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	64204	4.601	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	92.000%	
66) 1,2-Dichlorobenzene-d4	11.616	152	100384	4.940	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.800%	
Target Compounds						
					Qvalue	
13) Acetone	2.175	43	46281	26.585	ug/L	82
16) Methylene chloride	2.507	84	24520	1.181	ug/L	92
18) trans-1,2-Dichloroethene	2.764	96	6905	0.371	ug/L	85
22) cis-1,2-Dichloroethene	3.905	96	108953	5.756	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111422\
Data File : VV028929.D
Acq On : 14 Nov 2022 11:30
Operator : SY/MD
Sample : N5590-04DL 4X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBMX5DL

Quant Time: Nov 15 00:03:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 15 00:01:45 2022
Response via : Initial Calibration

