

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111422\
 Data File : VW028942.D
 Acq On : 14 Nov 2022 16:47
 Operator : SY/MD
 Sample : N5592-14DL 4X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AB9DL

Quant Time: Nov 15 00:05:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 15 00:01:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	268703	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	248826	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	116780	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	61124	3.731	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	74.600%	
7) Chloroethane-d5	1.568	69	61229	4.252	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	85.000%	
11) 1,1-Dichloroethene-d2	2.108	63	95000	3.182	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	63.600%	
20) 2-Butanone-d5	3.886	46	149700	54.591	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	109.180%	
24) Chloroform-d	4.342	84	153685	4.570	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.400%	
26) 1,2-Dichloroethane-d4	5.030	65	74073	4.895	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.000%	
32) Benzene-d6	5.047	84	321050	4.323	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.400%	
36) 1,2-Dichloropropane-d6	6.066	67	94795	4.571	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	91.400%	
41) Toluene-d8	7.313	98	283287	4.237	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.800%	
43) trans-1,3-Dichloroprop...	7.622	79	29226	4.321	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.400%	
46) 2-Hexanone-d5	8.085	63	127809	44.267	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	88.540%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	63402	4.409	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	88.200%	
66) 1,2-Dichlorobenzene-d4	11.619	152	99206	4.748	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	95.000%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.117	96	8887	0.546	ug/L #	1
16) Methylene chloride	2.506	84	27142	1.290	ug/L	95
19) 1,1-Dichloroethane	3.191	63	3094	0.100	ug/L #	82
22) cis-1,2-Dichloroethene	3.912	96	13744	0.717	ug/L	86
34) Trichloroethene	5.915	95	50932	2.616	ug/L	96
47) Tetrachloroethene	7.969	164	96342	6.181	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111422\
Data File : VV028942.D
Acq On : 14 Nov 2022 16:47
Operator : SY/MD
Sample : N5592-14DL 4X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AB9DL

Quant Time: Nov 15 00:05:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 15 00:01:45 2022
Response via : Initial Calibration

