

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111422\
 Data File : VW028947.D
 Acq On : 14 Nov 2022 18:50
 Operator : SY/MD
 Sample : N5592-19DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AC4DL

Quant Time: Nov 15 00:06:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 15 00:01:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	263875	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	245992	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	114309	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	61805	3.842	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.800%	
7) Chloroethane-d5	1.568	69	62970	4.453	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	89.000%	
11) 1,1-Dichloroethene-d2	2.108	63	93670	3.195	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	64.000%	
20) 2-Butanone-d5	3.895	46	159173	59.108	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.220%	
24) Chloroform-d	4.342	84	159738	4.836	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.800%	
26) 1,2-Dichloroethane-d4	5.027	65	74465	5.011	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.200%	
32) Benzene-d6	5.047	84	326498	4.447	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.000%	
36) 1,2-Dichloropropane-d6	6.066	67	97186	4.740	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	94.800%	
41) Toluene-d8	7.313	98	286091	4.328	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.600%	
43) trans-1,3-Dichloroprop...	7.622	79	29291	4.380	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.600%	
46) 2-Hexanone-d5	8.088	63	137075	48.023	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.040%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	69335	4.878	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.600%	
66) 1,2-Dichlorobenzene-d4	11.615	152	103008	5.037	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.800%	
Target Compounds						
12) 1,1-Dichloroethene	2.117	96	6158	0.385	ug/L #	1
16) Methylene chloride	2.503	84	37971	1.838	ug/L	93
22) cis-1,2-Dichloroethene	3.912	96	2278	0.121	ug/L	71
34) Trichloroethene	5.915	95	31011	1.611	ug/L	95
47) Tetrachloroethene	7.969	164	71267	4.625	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111422\
Data File : VV028947.D
Acq On : 14 Nov 2022 18:50
Operator : SY/MD
Sample : N5592-19DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AC4DL

Quant Time: Nov 15 00:06:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 15 00:01:45 2022
Response via : Initial Calibration

