

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111422\
 Data File : VW028948.D
 Acq On : 14 Nov 2022 19:14
 Operator : SY/MD
 Sample : N5592-21
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AE1

Quant Time: Nov 15 00:07:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 15 00:01:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	269219	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	249511	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	114034	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	60576	3.691	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	73.800%	
7) Chloroethane-d5	1.568	69	60861	4.218	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.400%	
11) 1,1-Dichloroethene-d2	2.108	63	90203	3.016	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	60.400%	
20) 2-Butanone-d5	3.886	46	152918	55.658	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	111.320%	
24) Chloroform-d	4.343	84	158892	4.715	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.400%	
26) 1,2-Dichloroethane-d4	5.027	65	75566	4.984	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.600%	
32) Benzene-d6	5.047	84	324531	4.358	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	87.200%	
36) 1,2-Dichloropropane-d6	6.066	67	97053	4.667	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.400%	
41) Toluene-d8	7.310	98	283044	4.222	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.400%	
43) trans-1,3-Dichloroprop...	7.622	79	28538	4.207	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.200%	
46) 2-Hexanone-d5	8.085	63	131649	45.472	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.940%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	70498	4.889	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.800%	
66) 1,2-Dichlorobenzene-d4	11.616	152	99674	4.886	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.800%	
Target Compounds						
						Qvalue
9) Trichlorofluoromethane	1.751	101	508607	17.755	ug/L	99
25) Chloroform	4.371	83	13866	0.428	ug/L	100
38) Bromodichloromethane	6.516	83	2421	0.109	ug/L	97
47) Tetrachloroethene	7.973	164	22842	1.461	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111422\
Data File : VV028948.D
Acq On : 14 Nov 2022 19:14
Operator : SY/MD
Sample : N5592-21
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AE1

Quant Time: Nov 15 00:07:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 15 00:01:45 2022
Response via : Initial Calibration

