

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111422\
 Data File : VW028958.D
 Acq On : 14 Nov 2022 23:43
 Operator : SY/MD
 Sample : N5604-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H46B9

Quant Time: Nov 15 05:35:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 15 01:23:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	268837	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	248951	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	125650	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	71928	4.389	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.800%
7) Chloroethane-d5	1.568	69	60180	4.177	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.600%
11) 1,1-Dichloroethene-d2	2.108	63	89713	3.004	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.000%
20) 2-Butanone-d5	3.889	46	69147	25.203	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	50.400%
24) Chloroform-d	4.342	84	152866	4.543	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.800%
26) 1,2-Dichloroethane-d4	5.027	65	68264	4.509	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
32) Benzene-d6	5.047	84	314396	4.231	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.600%
36) 1,2-Dichloropropane-d6	6.066	67	96252	4.639	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.800%
41) Toluene-d8	7.310	98	282979	4.230	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	7.622	79	24388	3.604	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.000%
46) 2-Hexanone-d5	8.088	63	121825	42.173	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.340%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	67549	4.695	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.000%
66) 1,2-Dichlorobenzene-d4	11.615	152	107680	4.790	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.800%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.310	62	1743021	82.158	ug/L	87
12) 1,1-Dichloroethene	2.117	96	15358	0.944	ug/L #	1
14) Carbon disulfide	2.294	76	27476	0.475	ug/L	98
18) trans-1,2-Dichloroethene	2.757	96	529769	28.067	ug/L	95
19) 1,1-Dichloroethane	3.188	63	12654	0.410	ug/L	95
22) cis-1,2-Dichloroethene	3.902	96	3185697	166.015	ug/L	96
33) Benzene	5.098	78	82997	1.091	ug/L	100
34) Trichloroethene	5.911	95	79427	4.078	ug/L	98
42) Toluene	7.387	91	77546	0.944	ug/L	98
51) Chlorobenzene	8.876	112	3003	0.059	ug/L	92
52) Ethylbenzene	9.011	91	36847	0.431	ug/L	97
53) m,p-Xylene	9.136	106	16229	0.477	ug/L	93
54) o-Xylene	9.542	106	3036	0.094	ug/L	89
60) Isopropylbenzene	9.931	105	5352	0.066	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111422\
Data File : VV028958.D
Acq On : 14 Nov 2022 23:43
Operator : SY/MD
Sample : N5604-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H46B9

Quant Time: Nov 15 05:35:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 15 01:23:16 2022
Response via : Initial Calibration

