

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111519\
 Data File : VV013643.D
 Acq On : 15 Nov 2019 17:03
 Operator : SY/MD
 Sample : K5855-11
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAQF9

Quant Time: Nov 19 01:23:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111519WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 23:49:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	375219	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	365255	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	165926	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	107322	4.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.60%
7) Chloroethane-d5	1.58	69	97670	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.40%
11) 1,1-Dichloroethene-d2	2.12	63	144158	3.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.40%
20) 2-Butanone-d5	3.96	46	250486	46.73	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.46%
24) Chloroform-d	4.39	84	232947	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.00%
26) 1,2-Dichloroethane-d4	5.08	65	115300	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	5.09	84	463653	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.00%
36) 1,2-Dichloropropane-d6	6.11	67	141382	4.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.00%
41) Toluene-d8	7.35	98	408156	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.20%
43) trans-1,3-Dichloropropene-	7.66	79	49999	4.32	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.40%
46) 2-Hexanone-d5	8.14	63	187404	42.77	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.54%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	92831	4.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	155112	4.86	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.20%

Target Compounds

					Ovalue
13) Acetone	2.23	43	13333	3.943	ug/L 82
42) Toluene	7.43	91	43000	0.364	ug/L 98
53) m,p-xylene	9.18	106	6220	0.131	ug/L 89
54) o-xylene	9.58	106	6656	0.146	ug/L 96
63) 1,4-Dichlorobenzene	11.31	146	24574	0.435	ug/L 96
68) 1,2,4-trichlorobenzene	13.31	180	5732	0.151	ug/L 89
70) 1,2,3-Trichlorobenzene	13.79	180	4017	0.116	ug/L # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111519\
Data File : VV013643.D
Acq On : 15 Nov 2019 17:03
Operator : SY/MD
Sample : K5855-11
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAQF9

Quant Time: Nov 19 01:23:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 23:49:05 2019
Response via : Initial Calibration

