


Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV111519\
 Data File : VV013634.D
 Acq On : 15 Nov 2019 08:40
 Operator : SY/MD
 Sample : BFB32
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB32

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111519WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri Nov 15 23:49:05 2019



AutoFind: Scans 2806, 2807, 2808; Background Corrected with Scan 2792

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	41170	PASS
75	95	30	80	48.2	114096	PASS
95	95	100	100	100.0	236842	PASS
96	95	5	9	6.8	16212	PASS
173	174	0.00	2	1.1	2386	PASS
174	95	50	120	95.9	227178	PASS
175	174	5	9	7.8	17717	PASS
176	174	95	101	96.5	219178	PASS
177	176	5	9	6.3	13703	PASS

m/z	Abundance
35.30	194.0
36.00	1991.0
37.10	10710.0
38.10	9347.0
39.00	3919.0
40.00	1222.0
41.10	181.0
42.10	176.0
43.00	271.0
44.00	1822.0
45.00	1999.0
46.00	348.0
47.10	2956.0
48.00	1433.0
49.10	8649.0
50.00	40592.0
51.00	12753.0
52.00	612.0
53.10	399.0
54.00	137.0
55.00	586.0
56.00	3065.0
57.00	5884.0
58.20	239.0
58.70	136.0
60.00	2224.0
61.00	10930.0
62.00	10678.0
63.00	8772.0
64.00	1681.0
65.00	1822.0
65.70	193.0
66.20	146.0
67.20	652.0
68.00	22832.0
69.10	22184.0
70.00	1808.0
71.00	118.0
72.10	1630.0
73.00	10784.0
74.10	38104.0
75.00	112960.0
76.10	10384.0
77.00	1567.0
78.00	895.0
79.00	6456.0
80.00	1954.0
80.90	6743.0
82.00	1440.0
82.80	164.0
86.00	623.0
86.90	8756.0
88.00	9489.0
90.90	741.0
92.00	6453.0
93.00	9932.0
94.10	27352.0
95.00	229824.0
96.10	16109.0
96.90	553.0
99.10	106.0
102.90	148.0
103.90	1030.0
105.00	201.0
105.90	995.0
107.00	285.0
108.20	167.0
109.80	199.0
111.00	338.0
111.80	134.0
113.00	189.0
114.90	320.0
115.90	701.0
117.00	1647.0
117.90	891.0
118.90	1506.0
122.30	201.0
123.10	136.0
123.80	189.0
127.00	120.0
127.90	847.0
128.80	371.0
129.90	894.0
130.90	380.0
134.90	367.0
136.90	464.0
138.20	107.0
139.10	179.0
139.70	141.0
141.00	2330.0
142.00	438.0
143.00	3526.0

Instrument :
MSVOA_V
ClientSampleId :
BFB32

144.00	208.0
144.90	1594.0
145.80	518.0
146.80	232.0
147.80	538.0
148.70	115.0
150.00	299.0
151.70	148.0
153.00	226.0
153.80	282.0
154.90	566.0
155.60	159.0
156.00	175.0
157.10	432.0
158.90	317.0
160.90	375.0
171.00	709.0
172.00	2910.0
173.00	3540.0
174.00	216256.0
175.00	17016.0
176.00	209216.0
177.00	13557.0
178.00	425.0
190.80	119.0
206.90	153.0
210.80	129.0
218.90	164.0
269.10	109.0

Instrument :
MSVOA_V
ClientSampleId :
BFB32

m/z	Abundance
35.20	174.0
36.00	1980.0
37.00	11375.0
38.10	9532.0
39.10	4046.0
40.00	1158.0
41.20	155.0
42.10	219.0
43.10	257.0
44.00	1864.0
45.00	2086.0
46.00	368.0
47.10	2836.0
48.00	1503.0
49.10	8816.0
50.00	42528.0
51.10	13184.0
52.00	525.0
53.10	459.0
54.20	136.0
55.00	632.0
56.10	3283.0
57.10	5883.0
58.20	305.0
58.80	130.0
60.00	2346.0
61.00	11045.0
62.00	10832.0
63.00	8726.0
64.00	1556.0
65.00	1894.0
66.20	165.0
67.00	705.0
68.00	23824.0
69.00	23408.0
70.10	1847.0
72.00	1664.0
73.00	10961.0
74.10	40640.0
75.00	118144.0
76.10	10698.0
77.00	1555.0
77.90	936.0
79.00	6901.0
80.00	2033.0
80.90	6743.0
82.00	1465.0
82.80	210.0
86.00	749.0
87.00	9468.0
88.00	10044.0
90.90	936.0
92.00	6918.0
93.00	10710.0
94.00	28848.0
95.00	244672.0
96.10	16832.0
96.90	544.0
97.30	336.0
99.10	128.0
102.90	136.0
103.90	1145.0
105.00	249.0
105.90	1046.0
106.90	256.0
108.20	127.0
109.80	207.0
111.00	304.0
111.90	164.0
113.10	228.0
114.90	325.0
115.90	715.0
117.00	1818.0
117.90	920.0
118.80	1519.0
122.30	144.0
122.70	142.0
123.80	239.0
125.40	103.0
127.00	121.0
127.90	931.0
128.80	445.0
129.90	968.0
130.90	313.0
134.90	394.0
136.90	500.0
138.20	118.0
139.10	157.0
139.60	160.0
141.00	2389.0
141.90	592.0
143.00	3643.0

Instrument :
MSVOA_V
ClientSampleId :
BFB32

144.00	233.0
145.00	1435.0
145.90	434.0
146.80	235.0
147.90	641.0
148.80	126.0
149.00	119.0
150.00	339.0
151.80	221.0
152.90	313.0
153.80	308.0
154.80	509.0
156.00	228.0
157.10	463.0
158.90	412.0
160.90	440.0
169.80	123.0
171.00	695.0
172.00	3018.0
173.00	3619.0
174.00	235008.0
175.00	18384.0
176.00	227200.0
177.00	14128.0
178.00	445.0
190.80	124.0
206.90	161.0
218.90	140.0
249.80	140.0
252.20	142.0
256.30	117.0

Instrument :
MSVOA_V
ClientSampleId :
BFB32

m/z	Abundance
35.10	176.0
36.00	1899.0
37.00	10703.0
38.10	8942.0
39.00	4038.0
40.00	1149.0
40.80	164.0
42.10	200.0
43.10	218.0
44.00	1736.0
45.00	1906.0
46.00	379.0
47.10	2580.0
48.00	1497.0
49.10	8402.0
50.00	40392.0
51.10	12573.0
52.10	483.0
53.10	499.0
54.20	155.0
55.00	624.0
56.00	3230.0
57.10	5482.0
58.20	266.0
58.80	106.0
60.00	2156.0
61.00	10250.0
62.00	10137.0
63.10	8264.0
64.00	1312.0
65.00	1856.0
66.10	210.0
67.00	701.0
68.00	22664.0
69.00	22312.0
70.10	1709.0
72.00	1387.0
73.00	10039.0
74.10	38608.0
75.10	111184.0
76.10	9588.0
77.00	1555.0
77.90	1019.0
79.00	6637.0
80.00	1924.0
80.90	6205.0
82.00	1411.0
82.80	227.0
85.10	123.0
86.00	697.0
87.00	9683.0
88.00	9430.0
90.90	945.0
92.00	6701.0
93.00	10015.0
94.00	27416.0
95.00	236032.0
96.10	15697.0
96.90	452.0
99.10	103.0
101.90	125.0
102.80	154.0
104.00	1034.0
104.90	383.0
106.00	946.0
106.90	241.0
109.70	185.0
111.00	238.0
112.00	195.0
113.10	192.0
114.80	307.0
115.90	714.0
117.00	1707.0
117.80	867.0
118.80	1333.0
122.70	133.0
123.90	243.0
124.90	127.0
125.30	106.0
126.00	118.0
127.90	1059.0
128.90	472.0
129.90	898.0
130.80	278.0
134.00	179.0
134.90	466.0
136.00	112.0
136.90	421.0
139.00	130.0
139.70	189.0
140.90	2220.0
141.90	581.0

Instrument :
MSVOA_V
ClientSampleId :
BFB32

142.90	3724.0
144.10	249.0
144.90	1187.0
146.00	402.0
146.80	234.0
147.30	171.0
147.90	656.0
149.00	170.0
149.90	371.0
151.80	259.0
152.90	308.0
153.80	293.0
154.80	489.0
156.00	193.0
157.00	457.0
157.80	128.0
159.00	379.0
160.80	434.0
169.80	161.0
171.00	587.0
172.00	3073.0
174.00	230272.0
175.00	17752.0
176.00	221120.0
177.00	13425.0
178.00	432.0
206.90	125.0
249.80	130.0
252.30	123.0
256.30	105.0
258.70	107.0
272.80	106.0

Instrument :
MSVOA_V
ClientSampleId :
BFB32

m/z	Abundance
35.20	129.0
35.90	227.0
36.40	176.0
37.00	263.0
38.00	192.0
38.70	141.0
39.90	739.0
40.90	175.0
42.00	150.0
42.70	144.0
43.90	918.0
47.20	159.0
48.00	107.0
49.20	129.0
51.60	100.0
52.20	151.0
52.90	141.0
54.00	114.0
54.90	208.0
56.10	106.0
59.30	195.0
60.80	116.0
62.30	124.0
63.40	101.0
65.80	113.0
67.80	118.0
69.10	123.0
77.10	155.0
77.80	129.0
78.50	117.0
80.60	158.0
84.00	107.0
87.70	118.0
88.80	130.0
91.10	133.0
97.10	106.0
111.60	106.0
122.30	149.0
127.10	117.0
145.10	108.0
150.40	146.0
205.90	105.0
210.80	104.0
222.90	101.0

Instrument :
MSVOA_V
ClientSampleId :
BFB32