

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111521\
 Data File : VV023482.D
 Acq On : 15 Nov 2021 15:07
 Operator : SY/MD
 Sample : M4616-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG1Z0

Quant Time: Nov 16 00:32:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 16 00:29:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	120056	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	118564	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	54424	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	41360	5.499	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 110.000%		
7) Chloroethane-d5	1.568	69	33560	5.475	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 109.400%		
11) 1,1-Dichloroethene-d2	2.108	63	56661	4.024	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 80.400%		
20) 2-Butanone-d5	3.896	46	69363	53.531	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 107.060%		
24) Chloroform-d	4.349	84	75263	4.696	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 94.000%		
26) 1,2-Dichloroethane-d4	5.034	65	35961	4.989	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 99.800%		
32) Benzene-d6	5.053	84	148429	4.879	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 97.600%		
36) 1,2-Dichloropropane-d6	6.069	67	44880	5.012	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 100.200%		
41) Toluene-d8	7.317	98	128034	4.491	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 89.800%		
43) trans-1,3-Dichloroprop...	7.629	79	16112	4.745	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 94.800%		
46) 2-Hexanone-d5	8.088	63	43796	35.055	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 70.120%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	26241	4.075	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 81.400%		
66) 1,2-Dichlorobenzene-d4	11.625	152	48300	5.330	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 106.600%		
Target Compounds						Qvalue
9) Trichlorofluoromethane	1.754	101	1155	0.077	ug/L	85
17) Methyl tert-butyl Ether	2.770	73	1776	0.113	ug/L	94
22) cis-1,2-Dichloroethene	3.925	96	1516	0.179	ug/L #	94
34) Trichloroethene	5.915	95	58701	6.661	ug/L	98
47) Tetrachloroethene	7.976	164	19287	2.525	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111521\
Data File : VV023482.D
Acq On : 15 Nov 2021 15:07
Operator : SY/MD
Sample : M4616-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG1Z0

Quant Time: Nov 16 00:32:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 16 00:29:25 2021
Response via : Initial Calibration

