

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111521\
 Data File : VV023490.D
 Acq On : 15 Nov 2021 18:18
 Operator : SY/MD
 Sample : M4617-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG224

Quant Time: Nov 16 00:34:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 16 00:29:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	120399	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	121379	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	54113	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	39554	5.244	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	104.800%	
7) Chloroethane-d5	1.568	69	31743	5.164	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	103.200%	
11) 1,1-Dichloroethene-d2	2.108	63	52779	3.738	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	74.800%	
20) 2-Butanone-d5	3.902	46	67843	52.209	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	104.420%	
24) Chloroform-d	4.349	84	72439	4.507	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.200%	
26) 1,2-Dichloroethane-d4	5.034	65	35412	4.899	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.000%	
32) Benzene-d6	5.050	84	142717	4.583	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.600%	
36) 1,2-Dichloropropane-d6	6.069	67	42286	4.612	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.200%	
41) Toluene-d8	7.317	98	123219	4.222	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.400%	
43) trans-1,3-Dichloroprop...	7.625	79	15235	4.383	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.600%	
46) 2-Hexanone-d5	8.092	63	43800	34.245	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	68.500%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	25668	3.893	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	77.800%	
66) 1,2-Dichlorobenzene-d4	11.625	152	46114	5.118	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.400%	
Target Compounds						
9) Trichlorofluoromethane	1.754	101	989	0.066	ug/L	82
17) Methyl tert-butyl Ether	2.770	73	6295	0.398	ug/L	96
34) Trichloroethene	5.924	95	3070	0.340	ug/L	78
47) Tetrachloroethene	7.976	164	20382	2.607	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111521\
Data File : VV023490.D
Acq On : 15 Nov 2021 18:18
Operator : SY/MD
Sample : M4617-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG224

Quant Time: Nov 16 00:34:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 16 00:29:25 2021
Response via : Initial Calibration

