

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111521\
 Data File : VV023497.D
 Acq On : 15 Nov 2021 21:05
 Operator : SY/MD
 Sample : VV1115WBL03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK254

Quant Time: Nov 16 02:09:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 16 02:06:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	124214	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	123389	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56127	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	41234	5.299	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	106.000%	
7) Chloroethane-d5	1.568	69	34164	5.387	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	107.800%	
11) 1,1-Dichloroethene-d2	2.108	63	57682	3.960	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	79.200%	
20) 2-Butanone-d5	3.908	46	67425	50.294	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.580%	
24) Chloroform-d	4.352	84	76562	4.617	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.400%	
26) 1,2-Dichloroethane-d4	5.034	65	38125	5.112	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.200%	
32) Benzene-d6	5.050	84	154833	4.891	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.800%	
36) 1,2-Dichloropropane-d6	6.069	67	45702	4.904	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.000%	
41) Toluene-d8	7.317	98	133667	4.505	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.200%	
43) trans-1,3-Dichloroprop...	7.625	79	16039	4.539	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.800%	
46) 2-Hexanone-d5	8.091	63	45587	35.062	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	70.120%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	26802	3.999	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	80.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	49255	5.270	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.400%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111521\
Data File : VV023497.D
Acq On : 15 Nov 2021 21:05
Operator : SY/MD
Sample : VV1115WBL03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK254

Quant Time: Nov 16 02:09:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 16 02:06:43 2021
Response via : Initial Calibration

