

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111521\
 Data File : VW023508.D
 Acq On : 16 Nov 2021 01:31
 Operator : SY/MD
 Sample : M4617-18
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG222

Quant Time: Nov 16 05:17:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 16 02:06:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	124316	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	125924	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	55452	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	39907	5.124	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	102.400%	
7) Chloroethane-d5	1.571	69	33030	5.204	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.000%	
11) 1,1-Dichloroethene-d2	2.111	63	54640	3.748	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	75.000%	
20) 2-Butanone-d5	3.896	46	68459	51.023	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.040%	
24) Chloroform-d	4.349	84	76012	4.580	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.600%	
26) 1,2-Dichloroethane-d4	5.034	65	36256	4.858	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.200%	
32) Benzene-d6	5.053	84	145725	4.510	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.200%	
36) 1,2-Dichloropropane-d6	6.069	67	43677	4.592	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	91.800%	
41) Toluene-d8	7.317	98	126557	4.180	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.600%	
43) trans-1,3-Dichloroprop...	7.625	79	14752	4.090	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	81.800%	
46) 2-Hexanone-d5	8.092	63	45021	33.929	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	67.860%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	25652	3.750	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	75.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	49422	5.353	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.000%	
Target Compounds						Qvalue
9) Trichlorofluoromethane	1.754	101	1333	0.086	ug/L	93
25) Chloroform	4.378	83	7143	0.436	ug/L	99
34) Trichloroethene	5.921	95	4373	0.467	ug/L	88
47) Tetrachloroethene	7.982	164	1115	0.137	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111521\
Data File : VV023508.D
Acq On : 16 Nov 2021 01:31
Operator : SY/MD
Sample : M4617-18
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG222

Quant Time: Nov 16 05:17:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 16 02:06:43 2021
Response via : Initial Calibration

