

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111522\
 Data File : VW028982.D
 Acq On : 15 Nov 2022 12:28
 Operator : SY/MD
 Sample : N5593-10DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AD9DL

Quant Time: Nov 16 03:17:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 16 03:16:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	281641	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	255798	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	118475	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	54907	3.198	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.000%
7) Chloroethane-d5	1.568	69	58855	3.899	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.000%
11) 1,1-Dichloroethene-d2	2.104	63	85882	2.745	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.800%#
20) 2-Butanone-d5	3.895	46	164953	57.391	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.780%
24) Chloroform-d	4.342	84	158615	4.499	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.000%
26) 1,2-Dichloroethane-d4	5.030	65	75077	4.733	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.600%
32) Benzene-d6	5.047	84	314836	4.123	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.400%
36) 1,2-Dichloropropane-d6	6.066	67	97350	4.566	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.400%
41) Toluene-d8	7.310	98	267684	3.895	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.800%
43) trans-1,3-Dichloroprop...	7.622	79	25122	3.613	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.200%
46) 2-Hexanone-d5	8.088	63	146757	49.444	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.880%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	72684	4.917	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.400%
66) 1,2-Dichlorobenzene-d4	11.615	152	101022	4.766	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.400%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.117	96	9291	0.545	ug/L #	1
16) Methylene chloride	2.503	84	30975	1.405	ug/L	94
22) cis-1,2-Dichloroethene	3.912	96	5086	0.253	ug/L	94
34) Trichloroethene	5.911	95	44822	2.240	ug/L	97
47) Tetrachloroethene	7.969	164	107471	6.707	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111522\
Data File : VV028982.D
Acq On : 15 Nov 2022 12:28
Operator : SY/MD
Sample : N5593-10DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AD9DL

Quant Time: Nov 16 03:17:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 16 03:16:14 2022
Response via : Initial Calibration

