

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111522\
 Data File : VW028984.D
 Acq On : 15 Nov 2022 13:17
 Operator : SY/MD
 Sample : N5593-12DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AE2DL

Quant Time: Nov 16 03:18:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 16 03:16:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	284609	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	256593	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	124603	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	54227	3.125	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	62.600%	
7) Chloroethane-d5	1.568	69	56516	3.705	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	74.200%	
11) 1,1-Dichloroethene-d2	2.105	63	87202	2.758	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	55.200%#	
20) 2-Butanone-d5	3.899	46	154184	53.084	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	106.160%	
24) Chloroform-d	4.342	84	156316	4.388	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	87.800%	
26) 1,2-Dichloroethane-d4	5.027	65	73948	4.614	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.200%	
32) Benzene-d6	5.043	84	314965	4.112	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.200%	
36) 1,2-Dichloropropane-d6	6.066	67	96234	4.500	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	90.000%	
41) Toluene-d8	7.310	98	268312	3.892	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.800%	
43) trans-1,3-Dichloroprop...	7.622	79	27156	3.893	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	77.800%	
46) 2-Hexanone-d5	8.088	63	143551	48.214	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.420%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	72384	4.882	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.600%	
66) 1,2-Dichlorobenzene-d4	11.615	152	100643	4.515	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	90.200%	
Target Compounds						
12) 1,1-Dichloroethene	2.114	96	11673	0.678	ug/L #	1
16) Methylene chloride	2.503	84	30925	1.388	ug/L	98
22) cis-1,2-Dichloroethene	3.915	96	5120	0.252	ug/L	93
34) Trichloroethene	5.911	95	65595	3.268	ug/L	98
47) Tetrachloroethene	7.969	164	146310	9.103	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111522\
Data File : VV028984.D
Acq On : 15 Nov 2022 13:17
Operator : SY/MD
Sample : N5593-12DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AE2DL

Quant Time: Nov 16 03:18:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 16 03:16:14 2022
Response via : Initial Calibration

