

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111522\
 Data File : VW029000.D
 Acq On : 15 Nov 2022 19:47
 Operator : SY/MD
 Sample : N5593-03DL 200X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AD1DL

Quant Time: Nov 16 03:22:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 16 03:16:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	260289	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	232142	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	111304	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	40219	2.535	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	50.600%	
7) Chloroethane-d5	1.561	69	47957	3.438	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	68.800%	
11) 1,1-Dichloroethene-d2	2.105	63	80456	2.782	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	55.600%#	
20) 2-Butanone-d5	3.915	46	145694	54.848	ug/L	0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	=	109.700%	
24) Chloroform-d	4.342	84	145503	4.466	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.400%	
26) 1,2-Dichloroethane-d4	5.031	65	71410	4.872	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.400%	
32) Benzene-d6	5.047	84	287107	4.143	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.800%	
36) 1,2-Dichloropropane-d6	6.066	67	90130	4.658	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.200%	
41) Toluene-d8	7.313	98	241851	3.877	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.600%	
43) trans-1,3-Dichloroprop...	7.622	79	24035	3.809	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	76.200%	
46) 2-Hexanone-d5	8.092	63	133549	49.579	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.160%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	70415	5.249	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	105.000%	
66) 1,2-Dichlorobenzene-d4	11.615	152	94156	4.728	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.600%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.114	96	26099	1.656	ug/L #	77
16) Methylene chloride	2.503	84	42306	2.076	ug/L	96
25) Chloroform	4.368	83	7159	0.229	ug/L	92
29) 1,1,1-Trichloroethane	4.606	97	3161	0.116	ug/L #	81
34) Trichloroethene	5.908	95	86245	4.749	ug/L	98
47) Tetrachloroethene	7.969	164	345428	23.754	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111522\
Data File : VV029000.D
Acq On : 15 Nov 2022 19:47
Operator : SY/MD
Sample : N5593-03DL 200X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AD1DL

Quant Time: Nov 16 03:22:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 16 03:16:14 2022
Response via : Initial Calibration

