

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111522\
 Data File : VW028989.D
 Acq On : 15 Nov 2022 15:19
 Operator : SY/MD
 Sample : N5621-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGCC2

Quant Time: Nov 16 03:19:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 16 03:16:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	262520	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	244805	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	115390	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	49272	3.079	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	61.600%	
7) Chloroethane-d5	1.568	69	54247	3.856	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	77.200%	
11) 1,1-Dichloroethene-d2	2.108	63	75417	2.586	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	51.800%#	
20) 2-Butanone-d5	3.895	46	104169	38.882	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	77.760%	
24) Chloroform-d	4.342	84	150439	4.578	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.600%	
26) 1,2-Dichloroethane-d4	5.030	65	71971	4.868	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.400%	
32) Benzene-d6	5.046	84	295685	4.047	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.000%	
36) 1,2-Dichloropropane-d6	6.066	67	92296	4.523	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	90.400%	
41) Toluene-d8	7.313	98	256969	3.907	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	78.200%	
43) trans-1,3-Dichloroprop...	7.625	79	26328	3.956	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	79.200%	
46) 2-Hexanone-d5	8.088	63	140935	49.615	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.220%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	72989	5.159	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.200%	
66) 1,2-Dichlorobenzene-d4	11.619	152	99828	4.836	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	96.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111522\
Data File : VV028989.D
Acq On : 15 Nov 2022 15:19
Operator : SY/MD
Sample : N5621-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGCC2

Quant Time: Nov 16 03:19:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 16 03:16:14 2022
Response via : Initial Calibration

