

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV111618\
 Data File : VV008575.D
 Acq On : 16 Nov 2018 10:08
 Operator : SY/MD
 Sample : VV1116MBL01
 Misc : 5.00g/5.0mL/100uL/5.0 mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK64

Quant Time: Nov 16 23:28:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 16 23:18:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	279838	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	250769	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	100497	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	97150	53.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	106.68%
7) Chloroethane-d5	1.58	69	74025	59.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	118.40%
11) 1,1-Dichloroethene-d2	2.13	63	137078	42.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.48%
21) 2-Butanone-d5	3.93	46	182521	115.43	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	115.43%
24) Chloroform-d	4.40	84	221267	54.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.82%
26) 1,2-Dichloroethane-d4	5.08	65	147743	56.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.74%
32) Benzene-d6	5.10	84	462620	56.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	113.44%
36) 1,2-Dichloropropane-d6	6.12	67	148208	56.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	113.02%
41) Toluene-d8	7.36	98	408823	55.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.36%
43) trans-1,3-Dichloropropene-	7.67	79	70237	55.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	110.98%
47) 2-Hexanone-d5	8.14	63	121869	108.83	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.83%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	184981	55.29	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	110.58%
64) 1,2-Dichlorobenzene-d4	11.68	152	135279	58.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	117.98%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV111618\
Data File : VV008575.D
Acq On : 16 Nov 2018 10:08
Operator : SY/MD
Sample : VV1116MBL01
Misc : 5.00g/5.0mL/100uL/5.0 mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK64

Quant Time: Nov 16 23:28:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111518WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 16 23:18:19 2018
Response via : Initial Calibration

