

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111621\
 Data File : VW023541.D
 Acq On : 16 Nov 2021 18:01
 Operator : SY/MD
 Sample : M4643-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB8K2

Quant Time: Nov 17 00:54:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 17 00:48:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	127825	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	124791	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	55768	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	28712	3.586	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.800%
7) Chloroethane-d5	1.568	69	24839	3.806	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.200%
11) 1,1-Dichloroethene-d2	2.108	63	42295	2.821	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.400%#
20) 2-Butanone-d5	3.911	46	38402	27.836	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	55.680%
24) Chloroform-d	4.349	84	60382	3.538	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	70.800%
26) 1,2-Dichloroethane-d4	5.034	65	31680	4.128	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
32) Benzene-d6	5.053	84	125584	3.922	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.400%
36) 1,2-Dichloropropane-d6	6.069	67	37081	3.934	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.600%
41) Toluene-d8	7.316	98	104876	3.495	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.000%
43) trans-1,3-Dichloroprop...	7.625	79	14072	3.937	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.800%
46) 2-Hexanone-d5	8.095	63	39801	30.268	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	60.540%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	22473	3.315	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	66.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	41883	4.510	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111621\
Data File : VV023541.D
Acq On : 16 Nov 2021 18:01
Operator : SY/MD
Sample : M4643-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB8K2

Quant Time: Nov 17 00:54:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 17 00:48:57 2021
Response via : Initial Calibration

