

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111622\
 Data File : VW029011.D
 Acq On : 16 Nov 2022 12:52
 Operator : SY/MD
 Sample : N5620-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGC99

Quant Time: Nov 17 02:52:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 17 02:49:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	257697	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	234843	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	112678	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	38970	2.481	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	49.600%	
7) Chloroethane-d5	1.574	69	45337	3.283	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	65.600%	
11) 1,1-Dichloroethene-d2	2.111	63	57262	2.000	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	40.000%#	
20) 2-Butanone-d5	3.970	46	129135	49.103	ug/L	0.07
Spiked Amount	50.000	Range 40 - 130	Recovery	=	98.200%	
24) Chloroform-d	4.349	84	124338	3.855	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	77.000%	
26) 1,2-Dichloroethane-d4	5.037	65	61188	4.216	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.400%	
32) Benzene-d6	5.040	84	220794	3.150	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	63.000%#	
36) 1,2-Dichloropropane-d6	6.069	67	86096	4.399	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	88.000%	
41) Toluene-d8	7.320	98	225277	3.570	ug/L	0.01
Spiked Amount	5.000	Range 70 - 130	Recovery	=	71.400%	
43) trans-1,3-Dichloroprop...	7.632	79	22091	3.460	ug/L	0.01
Spiked Amount	5.000	Range 55 - 130	Recovery	=	69.200%	
46) 2-Hexanone-d5	8.111	63	131493	48.254	ug/L	0.03
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.500%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	65336	4.814	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.200%	
66) 1,2-Dichlorobenzene-d4	11.616	152	89147	4.422	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	88.400%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111622\
Data File : VV029011.D
Acq On : 16 Nov 2022 12:52
Operator : SY/MD
Sample : N5620-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGC99

Quant Time: Nov 17 02:52:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 17 02:49:50 2022
Response via : Initial Calibration

