

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111622\
 Data File : VW029046.D
 Acq On : 17 Nov 2022 03:27
 Operator : SY/MD
 Sample : N5592-03DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AA8DL

Quant Time: Nov 17 05:37:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 17 04:20:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	268736	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	247452	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	115211	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	70957	4.331	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.600%	
7) Chloroethane-d5	1.568	69	58824	4.084	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	81.600%	
11) 1,1-Dichloroethene-d2	2.108	63	98635	3.304	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	66.000%	
20) 2-Butanone-d5	3.902	46	103532	37.751	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	75.500%	
24) Chloroform-d	4.342	84	155590	4.626	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.600%	
26) 1,2-Dichloroethane-d4	5.031	65	71091	4.697	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.000%	
32) Benzene-d6	5.047	84	328776	4.451	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.000%	
36) 1,2-Dichloropropane-d6	6.066	67	94319	4.573	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	91.400%	
41) Toluene-d8	7.310	98	310819	4.675	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.400%	
43) trans-1,3-Dichloroprop...	7.622	79	30408	4.520	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.400%	
46) 2-Hexanone-d5	8.088	63	128255	44.668	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	89.340%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	68162	4.767	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	95.400%	
66) 1,2-Dichlorobenzene-d4	11.615	152	110604	5.366	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.400%	
Target Compounds						
12) 1,1-Dichloroethene	2.117	96	10349	0.636	ug/L	1
16) Methylene chloride	2.507	84	22789	1.083	ug/L	98
34) Trichloroethene	5.912	95	96092	4.964	ug/L	97
47) Tetrachloroethene	7.969	164	100477	6.482	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111622\
Data File : VV029046.D
Acq On : 17 Nov 2022 03:27
Operator : SY/MD
Sample : N5592-03DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AA8DL

Quant Time: Nov 17 05:37:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 17 04:20:13 2022
Response via : Initial Calibration

