

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111622\
 Data File : VW029051.D
 Acq On : 17 Nov 2022 05:29
 Operator : SY/MD
 Sample : N5592-12DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AB7DL

Quant Time: Nov 17 06:30:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 17 04:20:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	274411	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	250335	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	118911	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	71261	4.260	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	85.200%	
7) Chloroethane-d5	1.568	69	58350	3.967	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	79.400%	
11) 1,1-Dichloroethene-d2	2.104	63	100413	3.294	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	65.800%	
20) 2-Butanone-d5	3.905	46	110123	39.324	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	78.640%	
24) Chloroform-d	4.342	84	160157	4.663	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.200%	
26) 1,2-Dichloroethane-d4	5.030	65	73410	4.750	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.000%	
32) Benzene-d6	5.047	84	338607	4.532	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.600%	
36) 1,2-Dichloropropane-d6	6.066	67	95407	4.573	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	91.400%	
41) Toluene-d8	7.310	98	311755	4.635	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.600%	
43) trans-1,3-Dichloroprop...	7.622	79	30760	4.520	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.400%	
46) 2-Hexanone-d5	8.088	63	131774	45.365	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.720%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	68988	4.769	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	95.400%	
66) 1,2-Dichlorobenzene-d4	11.615	152	115484	5.428	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	108.600%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.117	96	10786	0.649	ug/L	# 1
16) Methylene chloride	2.503	84	29008	1.351	ug/L	93
19) 1,1-Dichloroethane	3.188	63	17447	0.554	ug/L	98
22) cis-1,2-Dichloroethene	3.908	96	31163	1.591	ug/L	99
34) Trichloroethene	5.911	95	53029	2.708	ug/L	98
47) Tetrachloroethene	7.969	164	81384	5.190	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111622\
Data File : VV029051.D
Acq On : 17 Nov 2022 05:29
Operator : SY/MD
Sample : N5592-12DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AB7DL

Quant Time: Nov 17 06:30:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 17 04:20:13 2022
Response via : Initial Calibration

