

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111622\
 Data File : VW029053.D
 Acq On : 17 Nov 2022 06:18
 Operator : SY/MD
 Sample : N5592-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AB1

Quant Time: Nov 17 06:47:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 17 04:20:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	276198	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	250183	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	116873	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	76361	4.535	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	90.600%	
7) Chloroethane-d5	1.568	69	62771	4.240	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.800%	
11) 1,1-Dichloroethene-d2	2.108	63	121894	3.972	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	79.400%	
20) 2-Butanone-d5	3.896	46	166966	59.236	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.480%	
24) Chloroform-d	4.343	84	170252	4.925	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.400%	
26) 1,2-Dichloroethane-d4	5.031	65	77340	4.972	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.400%	
32) Benzene-d6	5.047	84	361257	4.838	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.800%	
36) 1,2-Dichloropropane-d6	6.066	67	103077	4.943	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.800%	
41) Toluene-d8	7.310	98	336407	5.004	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.000%	
43) trans-1,3-Dichloroprop...	7.619	79	33313	4.898	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	98.000%	
46) 2-Hexanone-d5	8.088	63	137146	47.243	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.480%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	74948	5.184	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.600%	
66) 1,2-Dichlorobenzene-d4	11.616	152	121176	5.795	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	116.000%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.118	96	44073	2.636	ug/L	# 80
19) 1,1-Dichloroethane	3.191	63	7394	0.233	ug/L	89
22) cis-1,2-Dichloroethene	3.912	96	25951	1.316	ug/L	94
25) Chloroform	4.375	83	10928	0.329	ug/L	83
34) Trichloroethene	5.908	95	220719	11.277	ug/L	98
47) Tetrachloroethene	7.969	164	571175	36.446	ug/L	99
51) Chlorobenzene	8.879	112	13332	0.261	ug/L	94
65) 1,4-Dichlorobenzene	11.265	146	27513	0.768	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111622\
Data File : VV029053.D
Acq On : 17 Nov 2022 06:18
Operator : SY/MD
Sample : N5592-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AB1

Quant Time: Nov 17 06:47:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 17 04:20:13 2022
Response via : Initial Calibration

