

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111622\
 Data File : VW029010.D
 Acq On : 16 Nov 2022 12:28
 Operator : SY/MD
 Sample : N5620-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGC98

Quant Time: Nov 17 02:52:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 17 02:49:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	255182	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	233908	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	109747	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.297	65	38081	2.448	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	49.000%
7) Chloroethane-d5	1.558	69	44588	3.260	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	65.200%
11) 1,1-Dichloroethene-d2	2.095	63	65804	2.321	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	46.400%#
20) 2-Butanone-d5	3.953	46	127132	48.818	ug/L	0.06
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.640%
24) Chloroform-d	4.339	84	143056	4.479	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
26) 1,2-Dichloroethane-d4	5.030	65	68698	4.780	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.600%
32) Benzene-d6	5.043	84	277128	3.969	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.400%
36) 1,2-Dichloropropane-d6	6.066	67	85044	4.362	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.200%
41) Toluene-d8	7.310	98	232280	3.696	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.000%
43) trans-1,3-Dichloroprop...	7.622	79	22417	3.525	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.600%
46) 2-Hexanone-d5	8.091	63	136943	50.455	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.920%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	67190	4.971	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.400%
66) 1,2-Dichlorobenzene-d4	11.615	152	93360	4.755	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111622\
Data File : VV029010.D
Acq On : 16 Nov 2022 12:28
Operator : SY/MD
Sample : N5620-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGC98

Quant Time: Nov 17 02:52:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 17 02:49:50 2022
Response via : Initial Calibration

