

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV111718\
 Data File : VV008593.D
 Acq On : 16 Nov 2018 20:26
 Operator : SY/MD
 Sample : J5987-05
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COL31

Quant Time: Nov 17 05:27:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 17 04:38:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	154174	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	139292	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	61045	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	35480	5.18	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.60%
7) Chloroethane-d5	1.59	69	27218	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.60%
11) 1,1-Dichloroethene-d2	2.14	63	50676	3.65	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.00%
20) 2-Butanone-d5	3.96	46	126822	51.75	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.50%
24) Chloroform-d	4.41	84	88083	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
26) 1,2-Dichloroethane-d4	5.09	65	47517	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
32) Benzene-d6	5.10	84	181781	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
36) 1,2-Dichloropropane-d6	6.12	67	58205	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.36	98	164938	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
43) trans-1,3-Dichloropropene-	7.67	79	20010	4.48	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.60%
46) 2-Hexanone-d5	8.14	63	94447	46.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.30%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	39404	4.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	55819	5.06	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV111718\
Data File : VV008593.D
Acq On : 16 Nov 2018 20:26
Operator : SY/MD
Sample : J5987-05
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COL31

Quant Time: Nov 17 05:27:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 17 04:38:54 2018
Response via : Initial Calibration

