

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV111718\
 Data File : VV008598.D
 Acq On : 16 Nov 2018 22:32
 Operator : SY/MD
 Sample : J5988-02
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JF2

Quant Time: Nov 17 05:38:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 17 04:38:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	154022	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	138638	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	61301	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	31742	4.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.80%
7) Chloroethane-d5	1.58	69	26431	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.80%
11) 1,1-Dichloroethene-d2	2.13	63	47162	3.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.00%
20) 2-Butanone-d5	3.98	46	119704	48.90	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.80%
24) Chloroform-d	4.41	84	86943	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.00%
26) 1,2-Dichloroethane-d4	5.09	65	47353	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
32) Benzene-d6	5.10	84	176117	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.40%
36) 1,2-Dichloropropane-d6	6.12	67	56818	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.36	98	163169	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	7.67	79	19927	4.48	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.60%
46) 2-Hexanone-d5	8.14	63	97236	48.26	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.52%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	38157	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	55864	5.05	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.00%

Target Compounds

					Qvalue
19) 1,1-Dichloroethane	3.23	63	3258	0.166 ug/L	99
22) cis-1,2-Dichloroethene	3.96	96	45846	3.919 ug/L	98
34) Trichloroethene	5.96	95	114053	9.817 ug/L	98
47) Tetrachloroethene	8.02	164	5188	0.617 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV111718\
Data File : VV008598.D
Acq On : 16 Nov 2018 22:32
Operator : SY/MD
Sample : J5988-02
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0JF2

Quant Time: Nov 17 05:38:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 17 04:38:54 2018
Response via : Initial Calibration

