

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV111718\
 Data File : VV008604.D
 Acq On : 17 Nov 2018 01:03
 Operator : SY/MD
 Sample : J5947-05
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9C04

Quant Time: Nov 17 05:54:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 17 05:41:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	166694	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	149656	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	65149	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	34680	4.69	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.80%
7) Chloroethane-d5	1.59	69	27289	4.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.60%
11) 1,1-Dichloroethene-d2	2.13	63	50356	3.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.00%
20) 2-Butanone-d5	3.96	46	118103	44.58	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.16%
24) Chloroform-d	4.41	84	87979	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.00%
26) 1,2-Dichloroethane-d4	5.09	65	47485	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
32) Benzene-d6	5.10	84	179346	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
36) 1,2-Dichloropropane-d6	6.12	67	58089	4.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.20%
41) Toluene-d8	7.36	98	166617	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
43) trans-1,3-Dichloropropene-	7.67	79	19291	4.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.40%
46) 2-Hexanone-d5	8.14	63	93812	43.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.26%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	38781	4.41	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	55156	4.69	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.80%

Target Compounds					Qvalue
30) Cyclohexane	4.72	56	5489	0.261 ug/L	96
47) Tetrachloroethene	8.02	164	6853	0.755 ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV111718\
Data File : VV008604.D
Acq On : 17 Nov 2018 01:03
Operator : SY/MD
Sample : J5947-05
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9C04

Quant Time: Nov 17 05:54:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 17 05:41:19 2018
Response via : Initial Calibration

