

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111720\
 Data File : VV019383.D
 Acq On : 17 Nov 2020 12:03
 Operator : SY/MD
 Sample : L4749-23
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BF8

Quant Time: Nov 18 05:51:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 17 17:05:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	225636	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	225317	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	124323	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	74358	43.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.44%
7) Chloroethane-d5	1.58	69	62304	45.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.16%
11) 1,1-Dichloroethene-d2	2.12	63	120455	34.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.84%
21) 2-Butanone-d5	3.91	46	73581	83.89	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.89%
24) Chloroform-d	4.37	84	142874	43.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.84%
26) 1,2-Dichloroethane-d4	5.05	65	98719	44.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.00%
32) Benzene-d6	5.07	84	261941	45.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.16%
36) 1,2-Dichloropropane-d6	6.09	67	74825	46.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.78%
41) Toluene-d8	7.33	98	249329	43.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.80%
43) trans-1,3-Dichloropropene-	7.64	79	43495	42.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.50%
47) 2-Hexanone-d5	8.10	63	43670	72.19	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	72.19%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	85882	38.98	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.96%
64) 1,2-Dichlorobenzene-d4	11.64	152	115882	48.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111720\
Data File : VV019383.D
Acq On : 17 Nov 2020 12:03
Operator : SY/MD
Sample : L4749-23
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BF8

Quant Time: Nov 18 05:51:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 17 17:05:19 2020
Response via : Initial Calibration

