

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111720\
 Data File : VV019385.D
 Acq On : 17 Nov 2020 13:17
 Operator : SY/MD
 Sample : L4710-18
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Nov 18 05:56:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 17 17:05:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	218547	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	221646	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	120648	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	74863	45.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.88%
7) Chloroethane-d5	1.58	69	63400	47.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.72%
11) 1,1-Dichloroethene-d2	2.12	63	121219	35.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.52%
21) 2-Butanone-d5	3.91	46	77768	91.54	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.54%
24) Chloroform-d	4.38	84	146754	46.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.14%
26) 1,2-Dichloroethane-d4	5.06	65	102716	47.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.52%
32) Benzene-d6	5.07	84	266608	46.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.28%
36) 1,2-Dichloropropane-d6	6.09	67	75688	47.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.40%
41) Toluene-d8	7.33	98	253485	45.37	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.74%
43) trans-1,3-Dichloropropene-	7.64	79	45046	44.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.98%
47) 2-Hexanone-d5	8.11	63	49473	83.14	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.14%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	92902	42.87	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.74%
64) 1,2-Dichlorobenzene-d4	11.65	152	117428	50.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.90%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111720\
Data File : VV019385.D
Acq On : 17 Nov 2020 13:17
Operator : SY/MD
Sample : L4710-18
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Nov 18 05:56:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 17 17:05:19 2020
Response via : Initial Calibration

