

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV111721\  
 Data File : VV023571.D  
 Acq On : 17 Nov 2021 15:47  
 Operator : SY/MD  
 Sample : M4627-12  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 H4631

Quant Time: Nov 18 00:22:12 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR110421WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Nov 18 00:20:29 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|--------|----------|
| -----                         |        |       |          |          |        |          |
| Internal Standards            |        |       |          |          |        |          |
| 1) 1,4-Difluorobenzene        | 5.616  | 114   | 110551   | 5.000    | ug/L   | 0.00     |
| 28) Chlorobenzene-d5          | 8.853  | 117   | 108113   | 5.000    | ug/L   | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 49518    | 5.000    | ug/L   | 0.00     |
|                               |        |       |          |          |        |          |
| System Monitoring Compounds   |        |       |          |          |        |          |
| 4) Vinyl Chloride-d3          | 1.304  | 65    | 41438    | 5.983    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =      | 119.600% |
| 7) Chloroethane-d5            | 1.564  | 69    | 30575    | 5.417    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =      | 108.400% |
| 11) 1,1-Dichloroethene-d2     | 2.108  | 63    | 54769    | 4.224    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =      | 84.400%  |
| 20) 2-Butanone-d5             | 3.889  | 46    | 58063    | 48.663   | ug/L   | -0.02    |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =      | 97.320%  |
| 24) Chloroform-d              | 4.349  | 84    | 66046    | 4.475    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =      | 89.400%  |
| 26) 1,2-Dichloroethane-d4     | 5.030  | 65    | 31828    | 4.795    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =      | 96.000%  |
| 32) Benzene-d6                | 5.050  | 84    | 133351   | 4.807    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =      | 96.200%  |
| 36) 1,2-Dichloropropane-d6    | 6.069  | 67    | 37102    | 4.544    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =      | 90.800%  |
| 41) Toluene-d8                | 7.317  | 98    | 116792   | 4.493    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =      | 89.800%  |
| 43) trans-1,3-Dichloroprop... | 7.625  | 79    | 13728    | 4.434    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =      | 88.600%  |
| 46) 2-Hexanone-d5             | 8.088  | 63    | 39163    | 34.377   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =      | 68.760%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84    | 21761    | 3.706    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =      | 74.200%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.625 | 152   | 42803    | 5.191    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =      | 103.800% |
|                               |        |       |          |          |        |          |
| Target Compounds              |        |       |          |          | Qvalue |          |
| 13) Acetone                   | 2.179  | 43    | 6112     | 8.383    | ug/L   | 97       |
| -----                         |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV111721\  
Data File : VV023571.D  
Acq On : 17 Nov 2021 15:47  
Operator : SY/MD  
Sample : M4627-12  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
H4631

Quant Time: Nov 18 00:22:12 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR110421WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Thu Nov 18 00:20:29 2021  
Response via : Initial Calibration

