

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111721\
 Data File : VW023586.D
 Acq On : 17 Nov 2021 21:47
 Operator : SY/MD
 Sample : M4627-14
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4632

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/18/2021
 Supervised By : Mahesh Dadoda 11/18/2021

Quant Time: Nov 18 00:25:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 18 00:20:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	131870	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	130300	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58667	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	39693	4.805	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.000%
7) Chloroethane-d5	1.568	69	29198	4.337	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.800%
11) 1,1-Dichloroethene-d2	2.105	63	54060	3.496	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.000%
20) 2-Butanone-d5	3.915	46	50504m	35.485	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	70.980%
24) Chloroform-d	4.349	84	64914	3.687	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.800%
26) 1,2-Dichloroethane-d4	5.034	65	30482	3.850	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.000%
32) Benzene-d6	5.050	84	129269	3.867	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.400%
36) 1,2-Dichloropropane-d6	6.069	67	36936	3.753	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.000%
41) Toluene-d8	7.317	98	112449	3.589	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.800%
43) trans-1,3-Dichloroprop...	7.625	79	13139	3.521	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.400%
46) 2-Hexanone-d5	8.095	63	43002	31.319	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	62.640%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	23325	3.296	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	66.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	40603	4.156	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111721\
Data File : VV023586.D
Acq On : 17 Nov 2021 21:47
Operator : SY/MD
Sample : M4627-14
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4632

Manual Integrations
APPROVED

Reviewed By : John Carlone 11/18/2021
Supervised By : Mahesh Dadoda 11/18/2021

Quant Time: Nov 18 00:25:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 18 00:20:29 2021
Response via : Initial Calibration

