

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111722\
 Data File : VW029065.D
 Acq On : 17 Nov 2022 15:50
 Operator : SY/MD
 Sample : N5628-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COL15

Quant Time: Nov 18 01:43:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 18 01:37:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	261161	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	239959	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	111872	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	74580	4.684	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.600%
7) Chloroethane-d5	1.568	69	55451	3.962	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.200%
11) 1,1-Dichloroethene-d2	2.108	63	91117	3.140	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.800%
20) 2-Butanone-d5	3.889	46	140828	52.839	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.680%
24) Chloroform-d	4.346	84	174957	5.352	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.000%
26) 1,2-Dichloroethane-d4	5.030	65	72800	4.950	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.000%
32) Benzene-d6	5.047	84	325175	4.540	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.800%
36) 1,2-Dichloropropane-d6	6.066	67	93581	4.679	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.600%
41) Toluene-d8	7.313	98	299847	4.651	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.000%
43) trans-1,3-Dichloroprop...	7.625	79	22409	3.435	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.800%
46) 2-Hexanone-d5	8.088	63	126281	45.354	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.700%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	67463	4.865	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.400%
66) 1,2-Dichlorobenzene-d4	11.615	152	111495	5.571	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.400%
Target Compounds						
25) Chloroform	4.368	83	458041	14.576	ug/L	95
38) Bromodichloromethane	6.509	83	43313	2.023	ug/L	99
42) Toluene	7.394	91	27360	0.345	ug/L	96
49) Dibromochloromethane	8.243	129	1406	0.105	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111722\
Data File : VV029065.D
Acq On : 17 Nov 2022 15:50
Operator : SY/MD
Sample : N5628-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COL15

Quant Time: Nov 18 01:43:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 18 01:37:41 2022
Response via : Initial Calibration

