

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111722\
 Data File : VW029082.D
 Acq On : 17 Nov 2022 22:56
 Operator : SY/MD
 Sample : N5687-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGCC7

Quant Time: Nov 18 06:00:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 18 04:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.603	114	236356	5.000	ug/L	-0.01
28) Chlorobenzene-d5	8.841	117	211108	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	97684	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.297	65	50431	3.500	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	70.000%	
7) Chloroethane-d5	1.558	69	47175	3.724	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	74.400%	
11) 1,1-Dichloroethene-d2	2.095	63	75024	2.857	ug/L	-0.02
Spiked Amount	5.000	Range 60 - 125	Recovery	=	57.200%#	
20) 2-Butanone-d5	3.960	46	122658	50.852	ug/L	0.06
Spiked Amount	50.000	Range 40 - 130	Recovery	=	101.700%	
24) Chloroform-d	4.336	84	128386	4.340	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.800%	
26) 1,2-Dichloroethane-d4	5.024	65	66438	4.991	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.800%	
32) Benzene-d6	5.030	84	264473	4.197	ug/L	-0.02
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.000%	
36) 1,2-Dichloropropane-d6	6.056	67	59287	3.369	ug/L	-0.01
Spiked Amount	5.000	Range 60 - 140	Recovery	=	67.400%	
41) Toluene-d8	7.297	98	263083	4.638	ug/L	-0.02
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.800%	
43) trans-1,3-Dichloroprop...	7.609	79	24114	4.202	ug/L	-0.01
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.000%	
46) 2-Hexanone-d5	8.104	63	110877	45.264	ug/L	0.02
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.520%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	61259	5.021	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	100.400%	
66) 1,2-Dichlorobenzene-d4	11.615	152	94026	5.380	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111722\
Data File : VV029082.D
Acq On : 17 Nov 2022 22:56
Operator : SY/MD
Sample : N5687-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGCC7

Quant Time: Nov 18 06:00:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 18 04:35:43 2022
Response via : Initial Calibration

