

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111722\
 Data File : WV029095.D
 Acq On : 18 Nov 2022 04:13
 Operator : SY/MD
 Sample : N5648-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H46E0

Quant Time: Nov 18 06:03:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 18 04:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	261310	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	238047	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	111037	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	60956	3.826	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.600%	
7) Chloroethane-d5	1.568	69	51989	3.712	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	74.200%	
11) 1,1-Dichloroethene-d2	2.108	63	84587	2.914	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	58.200%#	
20) 2-Butanone-d5	3.886	46	163491	61.307	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	122.620%	
24) Chloroform-d	4.342	84	148536	4.541	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.800%	
26) 1,2-Dichloroethane-d4	5.031	65	70696	4.804	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.000%	
32) Benzene-d6	5.047	84	305682	4.302	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.000%	
36) 1,2-Dichloropropane-d6	6.066	67	88943	4.483	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	89.600%	
41) Toluene-d8	7.310	98	280590	4.387	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.800%	
43) trans-1,3-Dichloroprop...	7.622	79	27431	4.239	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.800%	
46) 2-Hexanone-d5	8.088	63	128951	46.685	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	93.360%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	69684	5.066	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.400%	
66) 1,2-Dichlorobenzene-d4	11.615	152	104455	5.258	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.200%	
Target Compounds						
3) Chloromethane	1.240	50	6703	0.371	ug/L	90
5) Vinyl chloride	1.310	62	201793	9.786	ug/L	91
12) 1,1-Dichloroethene	2.114	96	2749	0.174	ug/L #	1
17) Methyl tert-butyl Ether	2.764	73	628326	18.527	ug/L	99
18) trans-1,2-Dichloroethene	2.757	96	301989	16.460	ug/L	94
19) 1,1-Dichloroethane	3.188	63	33570	1.119	ug/L	97
22) cis-1,2-Dichloroethene	3.902	96	1102963	59.134	ug/L	94
33) Benzene	5.098	78	24725	0.340	ug/L	100
34) Trichloroethene	5.911	95	77640	4.169	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111722\
Data File : VV029095.D
Acq On : 18 Nov 2022 04:13
Operator : SY/MD
Sample : N5648-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H46E0

Quant Time: Nov 18 06:03:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 18 04:35:43 2022
Response via : Initial Calibration

