

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111725\
 Data File : VV039391.D
 Acq On : 17 Nov 2025 14:57
 Operator : SY/MD
 Sample : VV1117MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK272

Quant Time: Nov 18 01:38:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111725WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 18 01:21:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	1107302	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.773	117	1131599	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.172	152	524512	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.291	65	513781	47.610	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.220%
7) Chloroethane-d5	1.545	69	387425	48.086	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.180%
11) 1,1-Dichloroethene-d2	2.072	65	215907	47.534	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.060%
21) 2-Butanone-d5	3.786	46	470728	82.613	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.610%
24) Chloroform-d	4.252	84	821030	45.907	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.820%
26) 1,2-Dichloroethane-d4	4.940	65	526133	47.274	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.540%
32) Benzene-d6	4.960	84	1624013	48.115	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.220%
36) 1,2-Dichloropropane-d6	5.982	67	503436	47.791	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.580%
41) Toluene-d8	7.236	98	1416876	45.452	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.900%
43) trans-1,3-Dichloroprop...	7.545	79	233291	43.894	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.780%
47) 2-Hexanone-d5	8.014	63	351324	87.301	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.300%
56) 1,1,2,2-Tetrachloroeth...	10.140	84	687194	44.629	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.260%
66) 1,2-Dichlorobenzene-d4	11.548	152	537023	54.508	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.020%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111725\
Data File : VV039391.D
Acq On : 17 Nov 2025 14:57
Operator : SY/MD
Sample : VV1117MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK272

Quant Time: Nov 18 01:38:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111725WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 18 01:21:06 2025
Response via : Initial Calibration

