

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV111818\
 Data File : VV008631.D
 Acq On : 17 Nov 2018 16:03
 Operator : SY/MD
 Sample : J5990-07MS
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETO69MS

Quant Time: Nov 19 06:38:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Nov 19 02:55:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	116549	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	107828	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	45543	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25757	4.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.60%
7) Chloroethane-d5	1.59	69	22706	5.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.60%
11) 1,1-Dichloroethene-d2	2.14	63	54791	5.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	104.40%
20) 2-Butanone-d5	3.95	46	96910	52.31	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.62%
24) Chloroform-d	4.41	84	67875	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	5.09	65	37234	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.40%
32) Benzene-d6	5.10	84	137739	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	6.12	67	44604	4.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.40%
41) Toluene-d8	7.36	98	125073	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
43) trans-1,3-Dichloropropene-	7.67	79	14543	4.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.00%
46) 2-Hexanone-d5	8.14	63	72894	46.51	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.02%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	30341	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	41036	4.99	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.80%

Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.14	96	30934	5.73	ug/L	96
13) Acetone	2.20	43	6625	6.86	ug/L	92
33) Benzene	5.15	78	174355	5.10	ug/L	100
34) Trichloroethene	5.96	95	46203	5.11	ug/L	97
42) Toluene	7.43	91	180084	5.06	ug/L	99
51) Chlorobenzene	8.93	112	112567	5.12	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV111818\
Data File : VV008631.D
Acq On : 17 Nov 2018 16:03
Operator : SY/MD
Sample : J5990-07MS
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETO69MS

Quant Time: Nov 19 06:38:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Nov 19 02:55:31 2018
Response via : Initial Calibration

