

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111822\
 Data File : VV029114.D
 Acq On : 18 Nov 2022 13:35
 Operator : SY/MD
 Sample : N5687-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGCD4

Quant Time: Nov 18 23:33:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 18 04:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	279274	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	255714	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	118392	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	79393	4.663	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	93.200%	
7) Chloroethane-d5	1.568	69	69038	4.612	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	92.200%	
11) 1,1-Dichloroethene-d2	2.108	63	107150	3.453	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	69.000%	
20) 2-Butanone-d5	3.912	46	169046	59.313	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.620%	
24) Chloroform-d	4.343	84	171301	4.901	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.000%	
26) 1,2-Dichloroethane-d4	5.031	65	84369	5.364	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	107.200%	
32) Benzene-d6	5.047	84	351667	4.607	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.200%	
36) 1,2-Dichloropropane-d6	6.066	67	109009	5.115	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.200%	
41) Toluene-d8	7.310	98	309669	4.507	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.200%	
43) trans-1,3-Dichloroprop...	7.622	79	33582	4.831	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	96.600%	
46) 2-Hexanone-d5	8.092	63	142513	48.030	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.060%	
56) 1,1,2,2-Tetrachloroeth...	10.211	84	80277	5.433	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	108.600%	
66) 1,2-Dichlorobenzene-d4	11.616	152	114284	5.395	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	108.000%	
Target Compounds						Qvalue
3) Chloromethane	1.240	50	3878	0.201	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111822\
Data File : VV029114.D
Acq On : 18 Nov 2022 13:35
Operator : SY/MD
Sample : N5687-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGCD4

Quant Time: Nov 18 23:33:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 18 04:35:43 2022
Response via : Initial Calibration

