

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111822\
 Data File : VW029138.D
 Acq On : 19 Nov 2022 00:00
 Operator : SY/MD
 Sample : N5688-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGCG1

Quant Time: Nov 19 02:00:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 18 04:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	275317	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	252701	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	107888	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	69118	4.118	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	82.400%	
7) Chloroethane-d5	1.568	69	63754	4.321	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	86.400%	
11) 1,1-Dichloroethene-d2	2.108	63	97091	3.174	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	63.400%	
20) 2-Butanone-d5	3.889	46	205073	72.988	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	145.980%#	
24) Chloroform-d	4.342	84	166143	4.821	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.400%	
26) 1,2-Dichloroethane-d4	5.027	65	82486	5.320	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.400%	
32) Benzene-d6	5.047	84	335888	4.453	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.000%	
36) 1,2-Dichloropropane-d6	6.066	67	108068	5.131	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.600%	
41) Toluene-d8	7.310	98	295344	4.350	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.000%	
43) trans-1,3-Dichloroprop...	7.622	79	32130	4.677	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	93.600%	
46) 2-Hexanone-d5	8.088	63	160452	54.721	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	109.440%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	83480	5.717	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	114.400%	
66) 1,2-Dichlorobenzene-d4	11.615	152	108152	5.603	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	112.000%	
Target Compounds						Qvalue
3) Chloromethane	1.240	50	2577	0.135	ug/L #	82
13) Acetone	2.179	43	7848	4.342	ug/L	84
25) Chloroform	4.375	83	8733	0.264	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111822\
Data File : VV029138.D
Acq On : 19 Nov 2022 00:00
Operator : SY/MD
Sample : N5688-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGCG1

Quant Time: Nov 19 02:00:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 18 04:35:43 2022
Response via : Initial Calibration

