

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111822\
 Data File : VW029155.D
 Acq On : 19 Nov 2022 06:55
 Operator : SY/MD
 Sample : N5666-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHB33

Quant Time: Nov 20 00:49:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sun Nov 20 00:44:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	296544	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	271056	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	126292	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	65506	3.623	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.400%
7) Chloroethane-d5	1.568	69	61975	3.899	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.000%
11) 1,1-Dichloroethene-d2	2.108	63	91942	2.791	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.800%#
20) 2-Butanone-d5	3.895	46	218021	72.042	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	144.080%#
24) Chloroform-d	4.342	84	164308	4.427	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
26) 1,2-Dichloroethane-d4	5.027	65	80445	4.817	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
32) Benzene-d6	5.047	84	325691	4.026	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.600%
36) 1,2-Dichloropropane-d6	6.066	67	105561	4.672	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.400%
41) Toluene-d8	7.313	98	284122	3.901	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.000%
43) trans-1,3-Dichloroprop...	7.622	79	30892	4.192	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.800%
46) 2-Hexanone-d5	8.088	63	160014	50.876	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.760%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	81927	5.230	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.600%
66) 1,2-Dichlorobenzene-d4	11.615	152	114065	5.048	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111822\
Data File : VV029155.D
Acq On : 19 Nov 2022 06:55
Operator : SY/MD
Sample : N5666-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHB33

Quant Time: Nov 20 00:49:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sun Nov 20 00:44:38 2022
Response via : Initial Calibration

