

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111822\
 Data File : VW029156.D
 Acq On : 19 Nov 2022 07:19
 Operator : SY/MD
 Sample : N5666-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHB35

Quant Time: Nov 20 00:49:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sun Nov 20 00:44:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	293315	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	267889	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	116301	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	65869	3.684	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	73.600%	
7) Chloroethane-d5	1.567	69	63174	4.019	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	80.400%	
11) 1,1-Dichloroethene-d2	2.108	63	92808	2.848	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	57.000%#	
20) 2-Butanone-d5	3.892	46	209986	70.151	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	140.300%#	
24) Chloroform-d	4.342	84	165338	4.504	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.000%	
26) 1,2-Dichloroethane-d4	5.030	65	80055	4.846	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.000%	
32) Benzene-d6	5.046	84	330731	4.136	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.800%	
36) 1,2-Dichloropropane-d6	6.066	67	104620	4.686	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.800%	
41) Toluene-d8	7.310	98	285272	3.963	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.200%	
43) trans-1,3-Dichloroprop...	7.622	79	30977	4.254	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	85.000%	
46) 2-Hexanone-d5	8.088	63	155051	49.881	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.760%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	81816	5.285	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	105.800%	
66) 1,2-Dichlorobenzene-d4	11.615	152	108067	5.194	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.800%	
Target Compounds						
18) trans-1,2-Dichloroethene	2.760	96	1871	0.091	ug/L	93
19) 1,1-Dichloroethane	3.188	63	13757	0.409	ug/L	98
29) 1,1,1-Trichloroethane	4.603	97	2707	0.086	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111822\
Data File : VV029156.D
Acq On : 19 Nov 2022 07:19
Operator : SY/MD
Sample : N5666-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHB35

Quant Time: Nov 20 00:49:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sun Nov 20 00:44:38 2022
Response via : Initial Calibration

