

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111918\
 Data File : VV008655.D
 Acq On : 19 Nov 2018 17:22
 Operator : SY/MD
 Sample : J5995-11
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0FM2

Quant Time: Nov 20 09:34:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 20 06:56:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	158293	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	144215	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	64760	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	29127	4.14	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.80%
7) Chloroethane-d5	1.58	69	25287	4.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.40%
11) 1,1-Dichloroethene-d2	2.13	63	44460	3.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.40%
20) 2-Butanone-d5	3.97	46	111249	44.22	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.44%
24) Chloroform-d	4.41	84	84122	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	5.09	65	44360	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
32) Benzene-d6	5.10	84	166043	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
36) 1,2-Dichloropropane-d6	6.12	67	54199	4.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.40%
41) Toluene-d8	7.36	98	150847	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
43) trans-1,3-Dichloropropene-	7.67	79	17336	3.75	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.00%
46) 2-Hexanone-d5	8.14	63	87379	41.69	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.38%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	37121	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	52898	4.52	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.40%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	11457	1.144 ug/L	93
18) trans-1,2-Dichloroethene	2.79	96	5742	0.721 ug/L	88
22) cis-1,2-Dichloroethene	3.96	96	4767	0.396 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111918\
Data File : VV008655.D
Acq On : 19 Nov 2018 17:22
Operator : SY/MD
Sample : J5995-11
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0FM2

Quant Time: Nov 20 09:34:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 20 06:56:52 2018
Response via : Initial Calibration

