

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111918\
 Data File : VV008660.D
 Acq On : 19 Nov 2018 19:28
 Operator : SY/MD
 Sample : J5995-16
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0FN8

Quant Time: Nov 20 08:55:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 20 06:56:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	152688	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	138660	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	59850	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	29227	4.31	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.20%
7) Chloroethane-d5	1.59	69	24710	4.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.40%
11) 1,1-Dichloroethene-d2	2.13	63	46109	3.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.00%
20) 2-Butanone-d5	3.96	46	118984	49.03	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.06%
24) Chloroform-d	4.41	84	84980	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
26) 1,2-Dichloroethane-d4	5.09	65	47146	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
32) Benzene-d6	5.10	84	165977	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
36) 1,2-Dichloropropane-d6	6.12	67	54846	4.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.00%
41) Toluene-d8	7.36	98	149966	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
43) trans-1,3-Dichloropropene-	7.67	79	17442	3.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.40%
46) 2-Hexanone-d5	8.14	63	88612	43.97	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.94%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	37744	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	51860	4.80	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.00%

Target Compounds					Ovalue
34) Trichloroethene	5.97	95	3292	0.283 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111918\
Data File : VV008660.D
Acq On : 19 Nov 2018 19:28
Operator : SY/MD
Sample : J5995-16
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0FN8

Quant Time: Nov 20 08:55:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 20 06:56:52 2018
Response via : Initial Calibration

