

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111921\
 Data File : VV023635.D
 Acq On : 19 Nov 2021 11:54
 Operator : SY/MD
 Sample : M4706-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
ClientSampleId :
 B0AB4

Manual Integrations
APPROVED
 Reviewed By :John Carlone 11/22/2021
 Supervised By :Mahesh Dadoda 11/22/2021

Quant Time: Nov 22 01:45:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 22 01:44:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	115178	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	115005	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	51629	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	31002	4.297	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.000%	
7) Chloroethane-d5	1.568	69	26810	4.559	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.200%	
11) 1,1-Dichloroethene-d2	2.108	63	45356	3.358	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	67.200%	
20) 2-Butanone-d5	3.908	46	53262m	42.846	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	85.700%	
24) Chloroform-d	4.352	84	64747	4.211	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.200%	
26) 1,2-Dichloroethane-d4	5.034	65	30487	4.409	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.200%	
32) Benzene-d6	5.053	84	117983	3.998	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	80.000%	
36) 1,2-Dichloropropane-d6	6.072	67	36989	4.258	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	85.200%	
41) Toluene-d8	7.317	98	100590	3.638	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	72.800%	
43) trans-1,3-Dichloroprop...	7.625	79	13237	4.019	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	80.400%	
46) 2-Hexanone-d5	8.091	63	43102	35.567	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	71.140%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	23795	3.809	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	76.200%	
66) 1,2-Dichlorobenzene-d4	11.625	152	40653	4.729	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.600%	
Target Compounds						
3) Chloromethane	1.240	50	3250	0.340	ug/L	94
13) Acetone	2.207	43	2264m	2.981	ug/L	
16) Methylene chloride	2.506	84	20098	2.005	ug/L	98
25) Chloroform	4.378	83	20403	1.343	ug/L	98
38) Bromodichloromethane	6.519	83	1597	0.159	ug/L #	93
48) 2-Hexanone	8.143	43	5221m	2.141	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111921\
 Data File : VV023635.D
 Acq On : 19 Nov 2021 11:54
 Operator : SY/MD
 Sample : M4706-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
ClientSampleId :
 B0AB4

Quant Time: Nov 22 01:45:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 22 01:44:25 2021
 Response via : Initial Calibration

Manual Integrations
APPROVED
 Reviewed By :John Carlone 11/22/2021
 Supervised By :Mahesh Dadoda 11/22/2021

