


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111921\
 Data File : VV023631.D
 Acq On : 19 Nov 2021 09:52
 Operator : SY/MD
 Sample : BFB238
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB238

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Mon Nov 22 01:44:25 2021



AutoFind: Scans 2809, 2810, 2811; Background Corrected with Scan 2797

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	13390	PASS
75	95	30	80	47.4	36995	PASS
95	95	100	100	100.0	77981	PASS
96	95	5	9	6.3	4948	PASS
173	174	0.00	2	0.9	647	PASS
174	95	50	120	91.5	71339	PASS
175	174	5	9	7.6	5443	PASS
176	174	95	105	96.0	68488	PASS
177	176	5	10	6.2	4257	PASS

m/z	Abundance
36.00	718.0
37.10	3387.0
38.00	2994.0
39.00	1285.0
39.90	121.0
44.00	456.0
45.00	560.0
46.20	191.0
47.00	894.0
47.90	456.0
49.10	2919.0
50.00	12757.0
51.00	4082.0
51.90	225.0
53.00	103.0
55.00	276.0
56.00	1040.0
57.00	2120.0
60.00	737.0
61.00	3610.0
62.00	3404.0
63.10	2973.0
64.00	421.0
65.00	605.0
65.90	141.0
66.90	253.0
68.00	7183.0
69.00	7847.0
70.00	627.0
72.10	456.0
73.00	3261.0
74.00	13540.0
75.10	35744.0
76.10	3040.0
76.90	416.0
78.00	244.0
78.90	2799.0
79.90	868.0
81.00	3064.0
81.90	832.0
87.00	2719.0
88.00	3035.0
90.90	352.0
92.00	2119.0
93.00	3205.0
94.00	9198.0
95.00	76272.0
96.10	4568.0
97.00	122.0
103.90	482.0
105.00	169.0
105.90	415.0
106.90	109.0
110.90	121.0
112.80	107.0
114.90	139.0
115.90	263.0
116.80	748.0
117.80	354.0
118.90	643.0
127.90	358.0
128.90	249.0
129.90	297.0
130.90	161.0
134.90	200.0
136.90	140.0
139.80	142.0
140.90	893.0
143.00	1145.0
144.90	425.0
146.00	158.0
146.80	111.0
147.90	309.0
150.00	103.0
155.00	233.0
156.90	164.0
161.00	117.0
171.00	202.0
172.00	978.0
174.00	70976.0
175.00	5078.0
176.00	68280.0
177.00	4237.0
178.00	132.0

Instrument :
MSVOA_V
ClientSampleId :
BFB238

m/z	Abundance
40.00	111.0
43.90	169.0

Instrument :
MSVOA_V
ClientSampleId :
BFB238

m/z	Abundance
36.00	625.0
37.00	3606.0
38.00	3127.0
39.00	1442.0
39.90	172.0
44.00	469.0
45.00	737.0
46.20	132.0
47.10	1009.0
48.00	413.0
49.10	3126.0
50.10	13644.0
51.00	4408.0
52.00	148.0
53.00	151.0
55.00	292.0
56.00	1179.0
57.00	2118.0
60.00	534.0
61.00	3766.0
62.10	3787.0
63.10	3107.0
64.00	468.0
65.00	566.0
65.90	128.0
66.90	171.0
68.00	7885.0
69.10	7893.0
70.00	570.0
72.00	532.0
73.00	3158.0
74.00	13599.0
75.10	37160.0
76.10	3299.0
77.10	350.0
77.90	313.0
78.90	2943.0
80.00	818.0
81.00	3236.0
82.00	844.0
87.00	2743.0
88.00	2821.0
90.90	415.0
92.00	2264.0
93.00	3405.0
94.00	9651.0
95.00	76816.0
96.00	5188.0
103.90	464.0
104.90	161.0
105.90	435.0
106.80	122.0
109.80	102.0
110.80	115.0
112.80	112.0
114.80	120.0
115.90	295.0
116.90	678.0
117.90	369.0
118.90	626.0
127.90	398.0
128.80	264.0
129.90	334.0
130.90	207.0
134.80	187.0
136.90	202.0
140.90	802.0
143.00	1315.0
144.80	364.0
145.90	159.0
147.90	329.0
150.00	121.0
154.90	255.0
157.00	193.0
159.00	123.0
170.90	155.0
171.90	1014.0
173.10	962.0
174.00	69192.0
175.00	5655.0
176.00	66296.0
177.00	4150.0
178.00	156.0

Instrument :
MSVOA_V
ClientSampleId :
BFB238

m/z	Abundance
36.00	695.0
37.10	3672.0
38.00	3148.0
39.00	1431.0
39.90	139.0
44.00	509.0
45.00	706.0
46.20	158.0
47.00	1005.0
48.00	458.0
49.10	3226.0
50.00	13768.0
51.00	4497.0
52.00	182.0
53.00	133.0
55.00	289.0
56.00	1115.0
57.00	2255.0
60.00	658.0
61.00	3846.0
62.00	3800.0
63.10	3247.0
64.00	463.0
65.00	588.0
65.90	175.0
66.90	215.0
68.00	7986.0
69.00	8316.0
70.00	615.0
72.00	479.0
73.00	3415.0
74.00	14165.0
75.10	38080.0
76.10	3332.0
77.00	362.0
77.90	274.0
78.90	2981.0
80.00	890.0
81.00	3325.0
81.90	874.0
87.00	2902.0
88.00	3008.0
90.90	411.0
92.00	2278.0
93.00	3511.0
94.00	9859.0
95.00	80856.0
96.00	5088.0
103.90	519.0
105.00	155.0
105.90	450.0
106.80	133.0
110.80	123.0
112.80	111.0
114.80	124.0
115.90	271.0
116.90	743.0
117.80	374.0
118.90	654.0
127.90	425.0
128.90	291.0
129.90	353.0
130.90	182.0
134.90	217.0
136.90	179.0
139.80	123.0
140.90	875.0
143.00	1257.0
144.90	413.0
146.00	161.0
147.90	345.0
150.00	103.0
154.90	244.0
157.00	185.0
159.10	123.0
170.90	183.0
172.00	1041.0
173.20	979.0
174.00	73848.0
175.00	5597.0
176.00	70888.0
177.00	4385.0
178.00	147.0

Instrument :
MSVOA_V
ClientSampleId :
BFB238