

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111922\
 Data File : VW029172.D
 Acq On : 19 Nov 2022 15:46
 Operator : SY/MD
 Sample : N5680-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AB9

Quant Time: Nov 20 00:58:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sun Nov 20 00:53:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	315657	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	292151	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	130909	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	106068	5.512	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 110.200%		
7) Chloroethane-d5	1.568	69	82569	4.881	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 97.600%		
11) 1,1-Dichloroethene-d2	2.105	63	129898	3.704	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 74.000%		
20) 2-Butanone-d5	3.899	46	307569	95.478	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 190.960%#		
24) Chloroform-d	4.342	84	187907	4.756	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 95.200%		
26) 1,2-Dichloroethane-d4	5.027	65	97453	5.482	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 109.600%		
32) Benzene-d6	5.047	84	388897	4.460	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 89.200%		
36) 1,2-Dichloropropane-d6	6.066	67	121318	4.982	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 99.600%		
41) Toluene-d8	7.310	98	336454	4.286	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 85.800%		
43) trans-1,3-Dichloroprop...	7.619	79	42476	5.348	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 107.000%		
46) 2-Hexanone-d5	8.085	63	285326	84.168	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 168.340%#		
56) 1,1,2,2-Tetrachloroeth...	10.207	84	108332	6.417	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 128.400%#		
66) 1,2-Dichlorobenzene-d4	11.615	152	132736	5.667	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 113.400%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111922\
Data File : VV029172.D
Acq On : 19 Nov 2022 15:46
Operator : SY/MD
Sample : N5680-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AB9

Quant Time: Nov 20 00:58:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sun Nov 20 00:53:03 2022
Response via : Initial Calibration

