

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111922\
 Data File : VW029197.D
 Acq On : 20 Nov 2022 02:09
 Operator : SY/MD
 Sample : N5648-15DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H46E2DL

Quant Time: Nov 20 04:31:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sun Nov 20 03:53:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	298559	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	277891	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	128781	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	93015	5.110	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	102.200%	
7) Chloroethane-d5	1.568	69	77486	4.843	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.800%	
11) 1,1-Dichloroethene-d2	2.105	63	117843	3.553	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	71.000%	
20) 2-Butanone-d5	3.905	46	113304	37.187	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	74.380%	
24) Chloroform-d	4.342	84	166525	4.456	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.200%	
26) 1,2-Dichloroethane-d4	5.030	65	79588	4.734	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.600%	
32) Benzene-d6	5.047	84	352768	4.253	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.000%	
36) 1,2-Dichloropropane-d6	6.066	67	103361	4.463	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	89.200%	
41) Toluene-d8	7.310	98	306250	4.102	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.000%	
43) trans-1,3-Dichloroprop...	7.622	79	30273	4.007	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	80.200%	
46) 2-Hexanone-d5	8.088	63	138687	43.010	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	86.020%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	74633	4.648	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	93.000%	
66) 1,2-Dichlorobenzene-d4	11.615	152	110433	4.793	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	95.800%	
Target Compounds						Qvalue
5) Vinyl chloride	1.310	62	11641	0.494	ug/L	87
16) Methylene chloride	2.506	84	5474	0.234	ug/L	97
17) Methyl tert-butyl Ether	2.764	73	359363	9.274	ug/L	99
18) trans-1,2-Dichloroethene	2.760	96	32294	1.541	ug/L	99
19) 1,1-Dichloroethane	3.188	63	2992	0.087	ug/L #	84
22) cis-1,2-Dichloroethene	3.908	96	128689	6.039	ug/L	99
34) Trichloroethene	5.918	95	6225	0.286	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111922\
Data File : VV029197.D
Acq On : 20 Nov 2022 02:09
Operator : SY/MD
Sample : N5648-15DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H46E2DL

Quant Time: Nov 20 04:31:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sun Nov 20 03:53:06 2022
Response via : Initial Calibration

