

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111922\
 Data File : VW029174.D
 Acq On : 19 Nov 2022 16:35
 Operator : SY/MD
 Sample : N5680-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AC1

Quant Time: Nov 20 00:59:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sun Nov 20 00:53:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	315901	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	294666	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	133785	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	103408	5.369	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 107.400%		
7) Chloroethane-d5	1.568	69	82456	4.870	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 97.400%		
11) 1,1-Dichloroethene-d2	2.108	63	127373	3.629	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 72.600%		
20) 2-Butanone-d5	3.899	46	402013	124.700	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 249.400%#		
24) Chloroform-d	4.343	84	187460	4.741	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 94.800%		
26) 1,2-Dichloroethane-d4	5.027	65	103528	5.819	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 116.400%		
32) Benzene-d6	5.043	84	384641	4.373	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 87.400%		
36) 1,2-Dichloropropane-d6	6.066	67	119741	4.875	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 97.600%		
41) Toluene-d8	7.310	98	327117	4.132	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 82.600%		
43) trans-1,3-Dichloroprop...	7.619	79	43909	5.482	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 109.600%		
46) 2-Hexanone-d5	8.085	63	381635	111.617	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 223.240%#		
56) 1,1,2,2-Tetrachloroeth...	10.207	84	121495	7.135	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 142.800%#		
66) 1,2-Dichlorobenzene-d4	11.616	152	130716	5.461	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 109.200%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111922\
Data File : VV029174.D
Acq On : 19 Nov 2022 16:35
Operator : SY/MD
Sample : N5680-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AC1

Quant Time: Nov 20 00:59:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sun Nov 20 00:53:03 2022
Response via : Initial Calibration

