

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW111922\  
 Data File : VW029175.D  
 Acq On : 19 Nov 2022 16:59  
 Operator : SY/MD  
 Sample : N5694-01  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 BHB37

Quant Time: Nov 20 00:59:27 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR111122WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Sun Nov 20 00:53:03 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc       | Units | Dev(Min) |
|-------------------------------|--------|----------------|----------|------------|-------|----------|
| -----                         |        |                |          |            |       |          |
| Internal Standards            |        |                |          |            |       |          |
| 1) 1,4-Difluorobenzene        | 5.613  | 114            | 276338   | 5.000      | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.847  | 117            | 259368   | 5.000      | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.243 | 152            | 115208   | 5.000      | ug/L  | 0.00     |
| System Monitoring Compounds   |        |                |          |            |       |          |
| 4) Vinyl Chloride-d3          | 1.307  | 65             | 105672   | 6.272      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | = 125.400% |       |          |
| 7) Chloroethane-d5            | 1.568  | 69             | 86767    | 5.859      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | = 117.200% |       |          |
| 11) 1,1-Dichloroethene-d2     | 2.108  | 63             | 134103   | 4.368      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | = 87.400%  |       |          |
| 20) 2-Butanone-d5             | 3.879  | 46             | 172949   | 61.327     | ug/L  | -0.02    |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | = 122.660% |       |          |
| 24) Chloroform-d              | 4.342  | 84             | 178853   | 5.171      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | = 103.400% |       |          |
| 26) 1,2-Dichloroethane-d4     | 5.031  | 65             | 83081    | 5.339      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | = 106.800% |       |          |
| 32) Benzene-d6                | 5.047  | 84             | 385242   | 4.976      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | = 99.600%  |       |          |
| 36) 1,2-Dichloropropane-d6    | 6.063  | 67             | 112522   | 5.205      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | = 104.000% |       |          |
| 41) Toluene-d8                | 7.310  | 98             | 332175   | 4.766      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | = 95.400%  |       |          |
| 43) trans-1,3-Dichloroprop... | 7.622  | 79             | 34543    | 4.899      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | = 98.000%  |       |          |
| 46) 2-Hexanone-d5             | 8.085  | 63             | 146906   | 48.813     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | = 97.620%  |       |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.207 | 84             | 77833    | 5.193      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | = 103.800% |       |          |
| 66) 1,2-Dichlorobenzene-d4    | 11.616 | 152            | 118300   | 5.739      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | = 114.800% |       |          |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV111922\  
Data File : VV029175.D  
Acq On : 19 Nov 2022 16:59  
Operator : SY/MD  
Sample : N5694-01  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
BHB37

Quant Time: Nov 20 00:59:27 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR111122WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Sun Nov 20 00:53:03 2022  
Response via : Initial Calibration

